UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/796,566	03/08/2004	Clark R. Baker JR.	TYHC:0069/FLE (P0426R)	1089
52144 Nellcor Puritan	7590 09/27/2007 Bennett LLC		EXAMINER	
c/o Fletcher Yoder PC			TOTH, KAREN E	
P.O. BOX 6922 HOUSTON, T			ART UNIT	PAPER NUMBER
·			3735	
			MAIL DATE	DELIVERY MODE
			09/27/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Application No.	Applicant(s)			
10/796,566	BAKER, CLARK R.	BAKER, CLARK R.		
Examiner	Art Unit			
Karen E. Toth	3735			

**Advisory Action** Before the Filing of an Appeal Brief --The MAILING DATE of this communication appears on the cover sheet with the correspondence address --THE REPLY FILED 31 August 2007 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. 1. The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods: months from the mailing date of the final rejection. a) The period for reply expres \_\_\_\_ b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is lateo. In r event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection. Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TW MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f). Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or set as the final office action action of the final office action action. above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). NOTICE OF APPEAL \_\_\_. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date 2. The Notice of Appeal was filed on \_ of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a). **AMENDMENTS** 3. The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because . (a) They raise new issues that would require further consideration and/or search (see NOTE below); (b) They raise the issue of new matter (see NOTE below); (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or (d) They present additional claims without canceling a corresponding number of finally rej ected claims. \_. (See 37 CFR 1.116 and 41.33(a)). 4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL -324). 5. Applicant's reply has overcome the following rejection(s): \_ 6. Newly proposed or amended claim(s) \_\_\_\_\_ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s). 7. Tor purposes of appeal, the proposed amendment(s): a) will not be entered, or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended. The status of the claim(s) is (or will be) as follows: Claim(s) allowed: Claim(s) objected to: Claim(s) rejected: Claim(s) withdrawn from consideration: \_\_\_ AFFIDAVIT OR OTHER EVIDENCE 8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e). 9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1). 10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached. REQUEST FOR RECONSIDERATION/OTHER 11. 🗵 The request for reconsideration has been considered but does NOT place the application in condition for allowance because: See Continuation Sheet. 12. Note the attached Information Disclosure Statement(s). (PTO/SB/08) Paper No(s). 13. Other: \_\_\_\_\_.

> CHARLES A. MARMOR II SUPERVISORY PATENT EXAMINER **TECHNOLOGY CENTER 3700**

Continuation of 11. does NOT place the application in condition for allowance because: Applicant argues that Diab (US 2002/0128544) does not disclose two methods of determining a heart rate; Diab clearly discloses separate methods of determining a heart rate - one in the case of no motion, and one in the case of motion (paragraphs [0327] -[0333]). Diab further discloses a selector for choosing a heart rate (analysis module - paragraphs [0332, [0345]) and evaluator for determining the reliability of a first h eart rate (High Confidence Test Module 570, which is directed to determining patient motion and may be used for either oxygen saturation or heart rate determination - see paragraph [0334]). Applicant's arguments are not persuasive, and the rejections stand as FINAL.